



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/620,312	07/14/2003	Michael P. Maher	AUOBIO.026CP1	7530
20995	7590	08/13/2007		
KNOBBE MARTENS OLSON & BEAR LLP 2040 MAIN STREET FOURTEENTH FLOOR IRVINE, CA 92614			EXAMINER PAK, MICHAEL D	
			ART UNIT 1646	PAPER NUMBER
			NOTIFICATION DATE 08/13/2007	DELIVERY MODE ELECTRONIC

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

jcartee@kmob.com
eOAPilot@kmob.com

Office Action Summary	Application No. 10/620,312	Applicant(s) MAHER ET AL.	
	Examiner Michael Pak	Art Unit 1646	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 31 May 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-10 is/are pending in the application.
- 4a) Of the above claim(s) 7-10 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-6 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Response to Amendment

1. Amendment filed May 31, 2007 has been entered.
2. Applicant's arguments filed May 31, 2007, have been fully considered but they are not found persuasive.
3. This application contains claims 7-10 drawn to an invention nonelected with traverse in the reply filed on August 2, 2006. A complete reply to the final rejection must include cancellation of nonelected claims or other appropriate action (37 CFR 1.144) See MPEP § 821.01.
4. Claims 1-6 are examined below.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

Art Unit: 1646

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

5. Claims 1-6 are provisionally rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 1-8, 10-29, and 49-50 of copending Application No. 09/804,457. Although the conflicting claims are not identical, they are not patentably distinct from each other.

Methods of claims 1-8, 10-29, and 49-50 of copending Application No. 09/804,457 anticipates or is obvious to use the method of claims 1-6 of the present application 10/620,312.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Applicants argue that pending claims of this application are patentable over this reference for at least the same reasons discussed below with reference to the applied art of Sinha. However, the double patenting is drawn to application 09/804,457 and examiner request that the arguments be specifically drawn to 09/804,457.

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the

Art Unit: 1646

art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

6. Claims 1-6 are rejected under 35 U.S.C. 112, first paragraph, because the specification, while being enabling for a method claimed where the field strength is 0.2% of the mean, does not reasonably provide enablement for a method claimed where the field strength is greater than 0.2% of the mean. The specification does not enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the invention commensurate in scope with these claims.

The first paragraph of § 112 requires that the patent specification enable "those skilled in the art how to make and use the full scope of the claimed invention without undue experimentation." Genentech, Inc. v. Novo Nordisk AIS, 108 F.3d 1361, 1365, 42 USPQ2d 1001, 1004 (Fed. Cir. 1997) (quoting In re Wright, 999 F.2d 1557, 1561, 27 USPQ2d 1510, 1513 (Fed. Cir. 1993)); see also In re Fisher, 427 F.2d 833, 839, 166 USPQ 18, 24 (CCPA 1970). ("[T]he scope of the claims must bear a reasonable correlation to the scope of enablement provided by the specification to persons of ordinary skill in the art."). Whether making and using the invention would have required undue experimentation, and thus whether the disclosure is enabling is a legal conclusion based upon several underlying factual inquiries. See In re Wands, 858 F.2d 731, 735, 736-37, 8 USPQ2d 1400, 1402, 1404 (Fed. Cir. 1988). As set forth in Wands, the factors to be considered in determining whether a claimed invention is enabled throughout its scope without undue experimentation include the quantity of experimentation necessary, the amount of direction or guidance presented, the

Art Unit: 1646

presence or absence of working examples, the nature of the invention, the state of the prior art, the relative skill of those in the art, the predictability or unpredictability of the art, and the breadth of the claims.

Likewise, in Amgen Inc. v. Chugai Pharm. Co., 927 F.2d 1200, 18 USPQ2d 1016 (Fed. Cir. 1991), the court affirmed the holding of invalidity of claims to analogs of the EPO gene under § 112 for lack of enablement where applicants had claimed every possible analog of the EPO gene but had disclosed only how to make EPO and a very few analogs. "[D]espite extensive statements in the specification concerning all analogs of the EPO gene that can be made, there is little enabling disclosure of the particular analogs and how to make them There may be many other genetic sequences that code for EPO-type products. Amgen has told how to make and use only a few of them and is therefore not entitled to claim all of them." *Id.*, 927 F.2d at 1213-14, 18 USPQ2d at 1027.

Claims encompass a method using electric field on cells in a well where the electric field strength can vary from the mean greater than 0.2%. However, one skilled in the art cannot make and use the method where the electric field strength is greater than 0.2% of the mean. The amount of direction provided in the specification is limited to using the method where the electric field strength is greater than 0.2% of the mean. However, the specification teaches that the electric field strength is greater than 0.2% of the mean is enabled (page 97 of the specification) whereas the electric field of 10% of the mean is not acceptable (page 96, line 10 of the specification). In view of the extent and the unpredictability of the experimentation required to practice the invention as

Art Unit: 1646

claimed, one skilled in the art could not make the invention without undue experimentation. Therefore, based on the above Wands analysis, a preponderance of the evidence supports a conclusion that one skilled in the art would not have been enabled to make and use the claimed invention without undue experimentation.

Applicants argue that page 95 of the specification provide enablement for "producing a limited spatial variation in intensity in [an] electric field in [an] area of observation of within the plus or minus 10% from a mean intensity in that area." However, it should be noted that the claims have not been amended to reflect the 10% from an mean intensity. Furthermore, page 96 specifically state that the geometry does not create uniform electrical fields and as a consequence is not suitable for use with the present invention. Thus the practice of the experimental invention is not suitable although there is a contradiction to the theoretical enablement of 10%. Page 95 of the specification clearly indicate that within the area of observation the standard deviation of the filed strength is 2% for figure 7A only. The breath of the claims do not reflect the enablement in the experimental part of the invention exemplified.

Priority

7. Applicant's claim for domestic priority under 35 U.S.C. 120 is acknowledged. However, the continuing application upon which priority is claimed fails to provide adequate support under 35 U.S.C. 112 for claims 1-6 of this application.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

The changes made to 35 U.S.C. 102(e) by the American Inventors Protection Act of 1999 (AIPA) and the Intellectual Property and High Technology Technical Amendments Act of 2002 do not apply when the reference is a U.S. patent resulting directly or indirectly from an international application filed before November 29, 2000. Therefore, the prior art date of the reference is determined under 35 U.S.C. 102(e) prior to the amendment by the AIPA (pre-AIPA 35 U.S.C. 102(e)).

8. Claims 1-6 are rejected under 35 U.S.C. 102(b) as being anticipated by Sinha et al. (1995).

Sinha et al. developed a system for simultaneous optical recording of transients of membrane potential and intracellular calcium concentration from mammalian brain slice preparations with high spatio-temporal resolution. Sinha et al. simultaneously record using two dedicated photodetectors together with two fluorescent indicators. The method uses hippocampal slice cells (page 53) exposed to glutamate antagonists (page 56 and figure 5) and cells are exposed to electric field. The cells are stimulated with 50 usec current pulses delivered at frequencies less than 0.05 Hz (page 53, column 2, last paragraph).

Art Unit: 1646

Applicants argue that there are not teach the pause between pulse series.

However, the specification does not specifically define the term "pulse series" nor distinguish that first and second pulse series be different. Since the pulse are in series and each pulse is separated by a pause there are a number of pulse series within the experiment which is separated by the frequency.

9. No claims are allowed.

10. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

11. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Pak whose telephone number is 571-272-0879. The examiner can normally be reached on 8:00 - 2:00.

Art Unit: 1646

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Gary Nickol can be reached on 571-272-0835. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Michael Pak
Primary Patent Examiner
Art Unit 1646
August 6, 2007